

# Instruction for Use KL02/IFU-E1

## Surgicare

## 01 Product Description

Product Name : Sterile Latex Surgical Gloves Powder Free

Polymer Coated

Material : Natural Rubber Latex Colour : White to Pale Yellow

Shape : Anatomic Cuff : Beaded

External Surface: Antitack Polymer treated

Internal Surface: Polymer Coated

Size : 5.5, 6.0, 6.5, 7.0, 7.5, 8.0, 8.5 and 9.0 Sterilization : Ethylene Oxide (EO) / GAMMA (R)

Shelf Life : 5 years

## 02 Intended Use

This disposal medical device is made up of natural rubber latex which is anatomically shaped with thumb position towards the Palm side of the index finger which reduces the fatigue on the hands, intended to be worn on the hands usually in surgical settings to provide barrier against potentially infectious fluids and other contaminants.

## 03 Standard compliance

EN 455-1:2020+A2:2024, EN 455-2:2024, EN 455-3:2023, EN 455-4:2009, EN ISO 15223-1:2021, EN 1041:2013, EN 566-1:2024, EN ISO 11135: 2014/A1:2019, EN ISO 11137-1:2015/A2:2019, EN ISO 11137-2:2015/A1:2023, EN ISO 11737-1:2018/A1:2021, EN ISO 11737-2:2020, ISO 11607-1:2020/A1:2023, ISO 11607-2:2020/A1:2023, EN ISO 10993-1:2020, EN ISO 10993-5:2009, EN ISO 10993-7:2008, EN ISO 10993-10:2021, EN ISO 10993-11:2017, EN ISO 10993-23:2021, ISO 10282:2023, ASTM D 3577:2019, ISO 13485:2016, ISO 14001:2015, ISO 9001:2015, IS 13422:2024.

## 04 Storage Instruction

Gloves must be stored in cool dry environment which is dust free. Cartons and Boxes must be stored unopened until required. Recommended storage temperature is 5°C-35°C. Avoid exposure to direct light, heat and excessive humidity.

As ozone is deleterious, storage area should not contain any equipment capable of generating ozone such as ultraviolet light, fluorescent lights, mercury vapour lamps, photocopier, high voltage equipment, x-ray units, electric motors and electro surgical equipments.

## 04 Storage Instruction

Gloves must be stored in cool dry environment which is dust free. Cartons and Boxes must be stored unopened until required. Recommended storage temperature is 5°C-35°C. Avoid exposure to direct light, heat and excessive humidity.

As ozone is deleterious, storage area should not contain any equipment capable of generating ozone such as ultraviolet light, fluorescent lights, mercury vapour lamps, photocopier, high voltage equipment, x-ray units, electric motors and electro surgical equipments.

## 05 Indication For Use

- Dry hands thoroughly before donning.
- Protective gloves should only be used for the intended application and in the correct size.
- These are sterile gloves for single use only.
- Users should take care when using the gloves. Using them solely according to their intended application.
- Before usage, inspect the gloves for any defect or imperfection.

## 06 Contraindication

- Latex gloves are made of Natural rubber latex, which may cause allergic reactions including anaphylaxis response if the user is allergic to latex.
- Gloves contain Natural Latex; persons who are sensitive to Latex should consult a physician before using.

## 07 Precautions

- Do not use if package is damaged or wet.
- Risk of reuse: May cause infection, allergic reaction and poor barrier protection.
- Gloves shall not be worn where there is a risk of entanglement by moving parts of machines is needed.
- The results do not reflect the actual duration of protection in the workplace due to other factors influencing the performance, such as temperature, abrasion, degradation etc.

## 08 Warnings

- Device disposal should be done as per local regulatory norms.
- Do not re-sterilize.
- The product contains Natural Rubber Latex which may cause allergic reactions including anaphylactic responses to some individuals.
- The gloves not intend to prevent Electrical shock care should be taken to have proper earthing in Medical Device Electrical appliances user.
- Necessary caution to be practiced against probable Electrical Hazards.




## Instruction for Use KL02/IFU-E1

## Surgicare


## Directions for use

## Glove Opening and Donning Procedure:



- Remove the walleted gloves (inner wrapper) from the Pouch (outer wrapper) by Peel open from the corner for Paper Pouch (Peel down to open pouch).
- © Open the walleted glove to see "Left" and "Right" compartment.
- Pinch back upper and lower flaps of the inner wrapper.
- Using the middle flaps, open the wrapper touching only the 1 inch margin for safety.
- Be sure wrapper does not close over gloves after opening to avoid contamination.
- " Using the thumb and the first two fingers of the nondominant hand, pinch the cuff of the folded edge of the glove cuff for the dominant hand, touching only the inside surface of the glove.
- Slide dominant hand in to the gloves keeping hand point downwards and pull up to wrist.
- Using the glove hand insert the 4 fingers under the cuff of the other glove and pull the glove up to the arm.
- Adjust the gloves as necessary.

## Glove Removal Procedure:



- Take hold of the first glove at the wrist.
- Fold it over and peel it back, turning it inside out as it goes. Once the glove is off, hold it with your gloved hand.
- To remove the other glove, place your bare fingers inside the cuff without touching the glove exterior. Peel the glove off from the inside, turning it inside out as it goes. Use it to envelope the other glove.

## **Explanation of Symbols**

Manufacturer

Date of Manufacture

Use by date

LOT

Lot No

SN

Serial Number

**STERILE** Sterilization using Ethylene oxide

**STERILE** R Sterilization using Irradiation



Do not re-sterilize



Do not use package is damaged or wet



Keep away from sunlight



Keep dry



Temperature limit



This way up or end up



Keep away from Ozone



Single Use



Instruction for Use



Caution



Latex Caution



# Instruction for Use KL02/IFU-E1

## Surgicare

## Personal Protective Equipment

EN ISO 374-1/Type A EN ISO 374-1/Type B EN ISO 374-1/Type C









TYPE P. Permeation min. 6 chemicals, level 2 (table 1)

TYPE B Permeation min. 3 chemicals, level 2 (table 1)

TYPE C Permeation min. 1 chemical, level 1 (table 1)

EN ISO 374-5:2016



Protective gloves protecting from viruses, bacteria and fungus

VIRUS



Protective gloves protecting from bacteria and fungus

Time

>10 min

>30 min

>60 min

>120 min

>240 min

#### Table 1

## Performance Level

Protection index

1

2

3

4

5

6

# Chemicals A Methanol B Acetone C Acetonitrile D Dichloromethane E Carbon disulphide F Toluene G Dimethylamine H Tetrahydrofuran

F	Toluene	>480 min
G	Dimethylamine	
Н	Tetrahydrofuran	
I	Ethyl acetate	

- J n-HeptaneK Sodium hydroxide 40%
- L Sulphuric acid 96%
- M Nitric acid 65%
- N Acetic acid 99%
- O Ammonium hydroxide 25%
- P Hydrogen peroxide 30%
- S Hydrofluoric acid 40%
- T Familia la 270
- T Formaldehyde 37%

#### EN ISO 374-1/Type B







**CE** 0.598

In Compliance with the Harmonized Standards Personal Protective Equipment, risk category III under Regulation (EU)2016/425

EN ISO 21420 : 2020 EN ISO 374-1 : 2016 EN ISO 374-2 : 2019 EN 16523-1 : 2015+A1:2018 EN ISO 374-4 : 2019 EN ISO 374-5 : 2016

Regulatory Authority : SGS Fimko Oy, Takomotie 8, FI-00380, Helsinki, Finland.

Notified Body Number : 0598

## 12 Warnings as per EN 374-1 & 5

- This information does not reflect the actual duration of protection in the work place and the differentiation between mixtures and pure chemicals.
- The Chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in cases where the glove is equal to or over 400mm where the cuff is tested also) and relates only to the chemical tested. It can be different if the chemical is used in a mixture.
- It is recommended to check that the gloves are suitable for the intended use because the conditions at the work place may differ from the type test depending on temperatures, abrasion and degradation.
- When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties.
- Movements, snagging, rubbing, degradation caused by the chemical contact etc., may reduce the actual use time significantly.
- For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves.
- Before usage, inspect the gloves for any defect or imperfections.
- For Single use only.
- The Penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen.

## Test Results as per EN ISO 374-1 & 5

## 1. EN ISO 21420: 2020

#### Protective Gloves - General Requirement and test methods

Туре	Declared Size	6.0	6.5	7.0	7.5	8.0	8.5
	Length (mm)	294.3	299.3	306.3	302	302.7	285.3
	Circumference (mm)	157.3	169.3	185.3	199.3	206.7	211.3
Glove	Dexterity	Performance Level 5					
	pH value	7.3					

#### Polyaromatic Hydrocarbons content

	Results
Benz (a) anthracene	Not Detected
Chrysene	Not Detected
Benzo (b) Fluoranthene	Not Detected
Benzo (k) Fluoranthene	Not Detected
Benzo (a) pyrene	Not Detected
Dibenz (a,h) anthracene	Not Detected
Benzo (e) pyrene	Not Detected
Benzo (j) Fluoranthene	Not Detected



# Instruction for Use KL02/IFU-E1

## Surgicare

## 2. EN ISO 374-2: 2019

Protective gloves against dangerous chemicals and micro-organisms - Part 2 : Determination of resistance to penetration

Resistance to Penetration	Status
Air Leakage	Pass
Water Leakage	Pass

Clause	Test Name	Test Results		Performance level
7.2	Air leak test (Air Pressure Used: 0.5 kPa)	Sample size 6.0 6.5 7.0 7.5 8.0	Leakage No Leakage No Leakage No Leakage No Leakage No Leakage	Pass
7.3	Water leak test	Sample size 6.0 6.5 7.0 7.5 8.0	Leakage No Leakage No Leakage No Leakage No Leakage No Leakage	Pass

## 3. EN ISO 374-4: 2019

Protective gloves against dangerous chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals

Chemical / CAS No	Exposure Duration	Test Results (Perchange in puncture	rcentage resistance	Observation
Sodium hydroxide 40% 1310-73-2	60±5 minutes	Glove sample 1 2 3 Mean Standard Deviation	Result (%) 6.9 7.3 6.6 6.9 0.330	No change
Sulphuric acid 96% 7664-93-9	60±5 minutes	Glove sample 1 2 3 Mean Standard Deviation	Result (%) 69.7 69.2 70.1 69.7 1.452	Severe Swelling & Colour change
Nitric acid 65% 7697-37-2	60±5 minutes	Glove sample 1 2 3 Mean Standard Deviation	Result (%) 68.1 67.5 67.1 67.6 0.505	Severe Swelling & Colour change
Acetic acid 99% 64-19-7	60±5 minutes	Glove sample 1 2 3 Mean Standard Deviation	Result (%) 27.3 28.6 27.3 27.7 0.761	Moderate swelling
Hydrogen peroxide 30% 7722-84-1	60±5 minutes	Glove sample 1 2 3 Mean Standard Deviation	Result (%) 6.3 7.2 6.6 6.7 0.500	Slight swelling
Formaldehyde 37% 50-00-0	60±5 minutes	Glove sample 1 2 3 Mean Standard Deviation	Result (%) 9.2 9.7 9.2 9.4 0.256	Slight swelling

## 4. EN 16523-1:2015+A1:2018

Determination of material resistance to permeation by chemicals - Part 1 : Permeation by Liquid chemical under conditions of continuous contact.

Chemical CAS No	Loop System /collection medium	Analytical Technique used	Mean thickness (mm)	NBT at NPR 1.0 µg cm <sup>-2</sup> min <sup>-1</sup> (minutes)	Performance level accordance to BS EN ISO 374-1:2016+ A1.2018 Table 1	Observation
Sodium hydroxide 40% 1310-73-2	Closed loop /Grade 3 water/350 rpm	Continuous measurement with conductivity electrode	0.21 0.20 0.21	>480 >480 >480 >480	6	No change
Nitric acid 65% 7697-37-2	Closed loop /Grade 3 water/350 rpm	Continuous measurement with conductivity electrode	0.21 0.20 0.21	53 50 56	2	Severe Swelling & Colour change
Formaldehyde 37% 50-00-0	Closed loop /Grade 3 water	Periodic measurement with HPLC	0.20 0.21 0.21	82 84 86	3	Slight Swelling
Sulphuric acid 96% 7664-93-9	Closed loop /Grade 3 water/350 rpm	Continuous measurement with conductivity electrode	0.20 0.20 0.21	17 19 22	1	Severe Swelling & Colour change
Acetic acid 99% 64-19-7	Closed loop /Grade 3 water/350 rpm	Continuous measurement with conductivity electrode	0.20 0.20 0.21	4 3 4	0	Moderate Swelling
Hydrogen peroxide 30% 7722-84-1	Closed loop /Grade 3 water	Periodic measurement with HPLC	0.20 0.21 0.21	5 6 7	0	Slight Swelling

## 5. EN ISO 374-5:2016

Protective gloves against dangerous chemicals and microorganisms - Part 5: Terminology and Performance Requirements for microorganisms risks

Protected against viruses according to ISO 16604 procedure B

Specimen	Thickness in mm	Observe the Presence of Plaques on Agar Plate (PFU)	Results (Pass/Fail)	Assay Titer Ratio
1	0.18	<1	Pass	
2	0.17	<1	Pass	
3	0.18	<1	Pass	1.3
Negative control		No Plaque formation	observed	
Positive control		Plaque formation obse		

Plate exposes during testing (Settle Plate): No plaque formation observed.

> Declaration of conformity is supplied with the product.